

NSLS Job Risk Assessment

The only official copy of this file is the one on-line in the NSLS ESH website. Before using a printed copy, verify that it is the most current version by checking the document issue date on the NSLS ESH website.

Name(s) of Risk Team Members: F. Zafonte, L. Stiegler, R. Greene	Point Value → Parameter ↓	1	2	3	4	5
Job Title: Radiation Generating Devices Job Number or Job Identifier: LS-JRA-0041	Frequency (B)	≤once/year	≤once/month	≤once/week	≤once/shift	>once/shift
Job Description: Use and operation of radiation generating devices (RGDs)	Severity (C)	First Aid Only	Medical Treatment	Lost Time	Partial Disability	Death or Permanent Disability
Training and Procedure List (Optional):	Likelihood (D)	Extremely Unlikely	Unlikely	Possible	Probable	Multiple
Approved by: A. Ackerman Date: 08/05/10 Rev. #: 2 Revision Log	Stressors (if applicable, please list all):		Reason for Revision (if applicable):		Comments:	

Job Step / Task	Hazard	Before Controls					Initial Controls	After Initial Controls					Control(s) Added to Reduce Risk	After Additional Controls				
		Stressors Y/N	# of People A	Frequency B	Severity C	Likelihood D		Risk* AxBxCxD	# of People A	Frequency B	Severity C	Likelihood D		Risk* AxBxCxD	# of People A	Frequency B	Severity C	Likelihood D

NSLS Job Risk Assessment

The only official copy of this file is the one on-line in the NSLS ESH website. Before using a printed copy, verify that it is the most current version by checking the document issue date on the NSLS ESH website.

Job Step / Task	Hazard	Before Controls						Initial Controls	After Initial Controls					Control(s) Added to Reduce Risk	After Additional Controls				
		Stressors Y/N	# of People A	Frequency B	Severity C	Likelihood D	Risk* AxBxCxD		# of People A	Frequency B	Severity C	Likelihood D	Risk* AxBxCxD		# of People A	Frequency B	Severity C	Likelihood D	Risk* AxBxCxD
Inserting or removing samples	Radiation exposure	N	1	2	2	3	12	Shielding, enclosure, interlocks, radiation postings, procedures, training, RGD Subject Area, Dosimetry, RadCon Manual	1	2	1	1	2						
Operation of device	Radiation exposure	N	1	2	3	3	18	Shielding, enclosure, interlocks, radiation postings, procedures, training, RGD Subject Area, Dosimetry, RadCon Manual	1	2	1	2	4						
	Electrical shock	N	1	2	5	3	30	LOTO, training,PPE, procedures, interlocks	1	2	2	2	8						

NSLS Job Risk Assessment

The only official copy of this file is the one on-line in the NSLS ESH website. Before using a printed copy, verify that it is the most current version by checking the document issue date on the NSLS ESH website.

Job Step / Task	Hazard	Before Controls						Initial Controls	After Initial Controls					Control(s) Added to Reduce Risk	After Additional Controls						
		Stressors Y/N	# of People A	Frequency B	Severity C	Likelihood D	Risk* AxBxCxD		# of People A	Frequency B	Severity C	Likelihood D	Risk* AxBxCxD		# of People A	Frequency B	Severity C	Likelihood D	Risk* AxBxCxD	% Risk Reduction	
Maintenance/ Alignment	Radiation exposure	N	1	2	3	3	18	Shielding, enclosure, interlocks, radiation postings, Dosimetry, procedures, training, RGD Subject Area, Work Planning and Control Documents	1	2	1	1	2								
	Electrical Shock	N	1	2	5	3	30	LOTO, training,PPE, procedures, interlocks	1	2	2	2	8								
Further Description of Controls Added to Reduce Risk:																					
*Risk:	0 to 20 Negligible	21 to 40 Acceptable					41 to 60 Moderate					61 to 80 Substantial					81 or greater Intolerable				