

Beamline X26A Hazardous Equipment List (rev. 11-08-2005)

Component	Energy Source	Specs	Quantity	Haz energies	Responsible Person	Location
Slit Tank	Ion pump controller Digital 500	5500V 350mA	1	Electrical	T Lanzirrotti and B Rao	In electronics rack next to slit tank
Slit Tank	Ion Gauge Controller J.C. Contolls	180V 25.5mA	1	Electrical	T Lanzirrotti and B Rao	In hutch on top of Alum. Shelf
Mirror Tank	Ion Pump Controller Digital 500	5500V 350mA	1	Electrical	T Lanzirrotti and B Rao	In electronics rack next to slit tank
Mirror Tank	Ion Gauge Controller HPS 919	180V 1mA	1	Electrical	T Lanzirrotti and B Rao	In hutch on top of Alum. Shelf
Canberra Detector	Iseg HV supply CHQ 102M	2kV 6mA	1	Electrical	T Lanzirrotti and B Rao	In electronics cabinet next to computers
Ion Chamber Inside hutch	Canberra HV supply 3002	3000V 3mA	1	Electrical	T Lanzirrotti and B Rao	In hutch on top of Alum. Shelf
Ion Chamber upstream of hutch wall	Canberra HV supply 3002	3000V 3mA	1	Electrical	T Lanzirrotti and B Rao	In hutch on top of Alum. Shelf
Vortex Detector	Radiant detector controller	500V 6mA	1	Electrical	T Lanzirrotti and B Rao	In hutch on top of Alum. Shelf