

APPENDIX G
USABILITY REPORTS

Attachment 1

BROOKHAVEN NATIONAL LABORATORY

Data Usability Summary Report

Project:	Sitewide RAV
Sampling Contractor:	R&C Formation
Analytical Laboratory:	Test America
Analytical Method:	905
Sample Delivery Group:	J3865-1
COC No. / Sample IDs:	33644-009
Date Sampled:	9/23/2013
Parameter(s):	Sr-90

<i>SAMPLE ID</i>	<i>DATA USABILITY QUALIFIERS</i>	<i>EXPLANATION FOR QUALIFIERS</i>
009	J-N2	Since the detection limit plus the error is greater than the value, there is evidence that the result may be a false positive. This combined with historic data results in a qualification of UJ-N2.

Summary:

Reviewed by: John Burke
Subject Matter Expert

Date: 1/30/2014

Reviewed by: _____
Project Manager

Date: _____