Accelerator Department
BROOKHAVEN NATIONAL LABORATORY
Associated Universities, Inc.
Upton, New York

AGS DIVISION TECHNICAL NOTE

No. 97

J. Guthy, J. Glenn, G.W. Bennett
July 26, 1972

G-10 90⁰ MONITOR TELESCOPE

This note updates AGS Tech Note No. 89 dated January 14, 1972.

The G-10 monitor telescope has been replated and recalibrated. We now read 1878 counts per 10^{12} protons. The high voltages are:

HVA = 2100

HVB = 1850

HVC = 2050

The change is partly due to recalibration of the circulating beam intensity monitor, aging and changes in temperature.

We expect to have this monitor recalibrated periodically.

Distr: Department Administration
AGS Division Staff
EP&S Division Staff
AGS Experimenters

BROOKHAVEN NATIONAL LABORATORY

BY *Anthony* DATE *7/26/77* SUBJECT *Cal. 10. Total*

CHKD. BY *C* DATE *7/27/77*

SHEET NO. 2 OF 2

JOB NO.

DEPT. OR PROJECT

SCINTILLATOR

DISC