A High Precision Instrument for Mapping of Rotational Errors in Rotary Stages

Authors

Weihe Xu^a, Kenneth Lauer^a, Yong S. Chu^a and Evgeny Nazaretski^{a*}

^aPhoton Sciences, Brookhaven National Laboratory, Upton, New York, 11973, USA

Correspondence email: enazaretski@bnl.gov

Synopsis A high-precision instrument suitable for mapping rotational errors in rotary stages has been developed. By performing comprehensive characterization of rotational errors, the performance of tomography systems can be drastically improved.

Abstract A rotational stage is a key component of every x-ray instrument capable of providing tomographic or diffraction measurements. To perform accurate 3D reconstructions run-out errors due to imperfect rotation (e.g. circle of confusion) must be quantified and corrected. We have developed a dedicated instrument capable of full characterization and circle of confusion mapping in rotary stages down to sub-10 nm level. A high stability design with an array of five capacitive sensors allows simultaneous measurements of wobble, radial and axial displacements. The developed instrument has been used for characterization of two mechanical stages which are part of an x-ray microscope.

1. Introduction

X-ray microscopy is a powerful tool suitable for studies of internal structure and composition of various material systems. [Ice, 2011, Kaulich *et al.*, 2011, Hyon Chol Kang *et al.*, 2013]. Recent developments of ultra-brilliant synchrotron facilities stimulated development of x-ray imaging techniques and pushed the imaging resolution down to nm-level [Yan *et al.*, 2014]. In the soft x-ray regime, 2D full-field imaging with better than 10 nm spatial resolution has been demonstrated using Fresnel zone plate as focusing optics [Chao *et al.*, 2012]. In the hard x-ray regime, X. Huang *et al.* reported 1D focusing down to 11 nm by Multilayers Laue Lenses (MLLs) [Huang *et al.*, 2013]. By taking a series of 2D x-ray projections at different angles a 3D image of a sample can be reconstructed revealing its internal structure and composition [Kaulich *et al.*, 2011, Holt *et al.*, 2013, Wu *et al.*, 2012]. For ptychography imaging, 16 nm spatial resolution has been reported in 3D. [Holler *et al.*, 2012 & 2014] For all tomographic measurements, particularly via scanning x-ray microscopy, rotational errors play a critical role in defining scientific throughput of a microscope unless actively corrected [Holler *et al.*, 2012 & 2014] Therefore, characterization of displacements due to wobble, axial and radial runouts in rotary stages prior to the measurements is necessary during development of microscopy systems for 3D nanotomography.

A typical method of characterizing rotary stages includes the attachment of a ball gauge to the center of a stage and monitoring its displacement by sensors when rotating [Noire *et al.*, 2010]. Non-contact displacement sensors (capacitive, optical *etc.*) are usually employed for high precision characterization [Marco, 2010 & Kim *et al.*, 2013]. Unfortunately, such ball gauge systems do not provide simultaneous measurements of displacement errors due to wobble, radial and axial runout errors. In this work, we present a system capable of comprehensive characterization of all rotational errors during the same measurement. By incorporating five capacitive sensors and a diamond turned reference cylinder nm-scale rotational errors can be measured. A mechanically stiff yet instrumentally flexible design, allows for mapping of rotational errors in stages with different form factors and footprints.

2. Instrument Design

A Computer Aided Design (CAD) drawing and a photograph of the instrument are shown in Fig. 1a and 1b. The instrument has a robust yet flexible frame structure to provide stability while allowing characterization of rotary stages with different footprints and form factors. Five Lion Precision CPL 190 capacitive displacement sensors with C5S-2.0 probes were arranged in the middle of the beam as shown in the lower inset of Fig 1b. Sensors S1 and S2 measure the displacement of the cylinder in the X-direction; S3 and S4 in the Y-direction and S5 in the Z-direction respectively. The rotary stage under investigation is placed on top of a high precision Kohzu YM16A-S1 X-Y manual linear stage. The linear stage provides 25 mm travel in both X and Y directions and yields 0.5 μm resolution. The manual stage is used to position rotary stage to be characterized within the working ranges of capacitive sensors. On top of the rotary stage, there is an assembly of X and Y linear piezo stages (Attocube ECS3030/NUM) with the diamond turned reference cylinder finally on top of that. Piezo stages are used for the alignment of the cylinder axis and the rotary stage center of rotation. The reference cylinder is 40 mm tall and 25 mm in diameter. It also has a diamond-turned top surface with a machined pin in the center and used for initial optical alignment with a CCD camera (BASLER Ace acA1300-30gc). The surface roughness of the reference cylinder was measured by a Zygo optical surface profiler and yielded peak to valley values of ± 11 nm (see Fig. 1c, 1d). Since the diameter of the capacitive C5S-2.0 probe is about 2 mm, the localized roughness is averaged and it provides an RMS noise of approximately ± 2 nm. The system is largely controlled through a LabviewTM interface. The interface communicates with some hardware directly and indirectly interfaces with the remainder through the Experimental Physics and Industrial Control System (EPICS) channel access protocol.

3. Testing Procedures and Experimental Result

When rotational motion is performed, lateral displacement of a cylinder D can be expressed as follows:

$$D = D_{\text{runout}} + D_{\text{wobble}} + D_{\text{off-center}}, \quad (1)$$

where D_{runout} and D_{wobble} are the displacements caused by radial runout and wobble. Some of the errors are synchronous (repeatable) and some of the errors are asynchronous (random). $D_{\text{off-center}}$ is the displacement caused by misalignment between the axis of a cylinder and rotational axis of the rotary stage. Displacements measured by sensors S1-S4 can be written as shown below:

$$D_1 = A \cos \alpha + L \sin \theta \cos \beta + R_{13} \cos \omega t \quad (2)$$

$$D_2 = A \cos \alpha + (L - d_{12}) \sin \theta \cos \beta + R_{24} \cos \omega t \quad (3)$$

$$D_3 = A \sin \alpha + L \sin \theta \sin \beta + R_{13} \sin \omega t \quad (4)$$

$$D_4 = A \sin \alpha + (L - d_{12}) \sin \theta \sin \beta + R_{24} \sin \omega t, \quad (5)$$

where A and α are the magnitude and phase of runout. θ and β are the magnitude and phase of wobble. L is the measured distance from S-1 to the top surface of the rotary stage. d_{12} is the distance between S-1 and S-2. R_{13} and R_{24} are the distances between the cylinder's center of rotation in the planes of S1-S3 and S2-S4. By fitting the measured values of D_{1-4} to $\sin \omega t$ and $\cos \omega t$, values of R_{13} and R_{24} can be calculated and consequently subtracted from D_{1-4} . Finally, since values of L and d_{12} are directly measured, values of θ and β can be calculated yielding A and α .

Displacements measured by S-5 can be written as follows:

$$\begin{aligned} D_5 &= D_{\text{wobble}} + D_{\text{tilt}} + D_{\text{vertical_displacement}} \\ &= R_5 \sin \theta \cos \beta + R_5 \sin \gamma \cos \omega t + D_{\text{vertical_displacement}}, \end{aligned} \quad (6)$$

where D_{tilt} is the displacement caused by tilting of the cylinder. $D_{\text{vertical_displacement}}$ is the displacement caused by the vertical motion (axial runout). R_5 is the distance from the measurement point of sensor S-5 to the rotation center. γ is the tilting angle of the cylinder. By fitting the measured values of D_5 to $\cos \omega t$, values of R_5 can be calculated. Since the values of θ and β at each rotational angle have been measured, $D_{\text{vertical_displacement}}$ can then be calculated.

To verify the performance of the developed system we fully characterized rotational errors of two mechanical rotary stages, referred to here as Stage 1 and Stage 2. Stage 1 is Attocube ANR240/RES. It has a small footprint, uses a stick-slip piezoelectric actuator and is encoded from 0° to 320° (shown in Fig. 1a, 1b). Stage 2 is a KOHZU RA10A-W. It has a larger form factor and has a rotation range between 0° and 280° . Measurements for each stage were repeated six times, and the results are depicted in Fig. 2. For Stage 1 radial runout, wobble and axial runout were measured be $6 \pm 5 \mu\text{m}$, $200 \pm 150 \mu\text{rad}$ and $\pm 0.6 \mu\text{m}$ respectively. For Stage 2, the measured values were $0.4 \pm 0.4 \mu\text{m}$, $6 \pm 6 \mu\text{rad}$ and $\pm 0.08 \mu\text{m}$ accordingly. Based on the measurements shown in Fig. 2 one can determine that runout and wobble for both stages are highly repeatable. The most prominent example is the 2 degree periodic error measured for Stage 2. Since the majority of rotational errors are reproducible,

synchronous deviations from a perfect trajectory can be programmed into tomographic data acquisition software, and the field of view can be adjusted accordingly during measurements of each individual projection. In doing so, the data acquisition time can be reduced yielding higher throughput of a tomographic system.

4. Conclusion

In this work, we have developed a high-precision instrument suitable for the characterization of rotational stages used in 3D tomographic measurements. The instrument provides simultaneous measurements of both radial and axial runouts and wobble in a rotary stage. Stable structural design, application of five capacitor sensors and diamond turned reference cylinder ensure nm-level mapping of rotational errors. The developed system was used to characterize two mechanical rotary stages and revealed both synchronous and asynchronous errors.

Acknowledgements

We acknowledge B. Mullany (BNL) for help with 3D modeling of the microscope and D. Kuhne (BNL) for machining/assembling of mechanical parts. Work at Brookhaven was supported by the US Department of Energy under Contract No. DE-AC02-98CH10886.

References

- Ice, G. E., Budai, J. D. & Pang, J. W. L. (2011). *Science*, 334, 1234-1239
- Kaulich B., Thibault P., Gianoncelli A., and Kiskinova M. (2011). *J. Phys.: Condens. Matter*, 23 83002
- Kang Hyon Chol, Yan Hanfei, Chu Yong S., Lee Su Yong, Kim Jungdae, Nazaretski Evgeny, Kim Chan, Seo Okkyun, Noh Do Young, Macrander Albert T., Stephenson G. Brian and Maser Jorg (2013). *Nanoscale*, 5 7184-7187
- Chao, W., Fischer, P., Tyliczszak, T., Rekawa, S., Anderson, E. & Naulleau, P. (2012). *Opt. Express*, 20, 9777-9783
- Huang, X., Yan, H., Nazaretski, E., Conley, R., Bouet, N., Zhou, J., Lauer, K., Li, L., Eom, D., Legnini, D., Harder, R., Robison, I. K. & Chu, Y-S. (2013). *Review of Scientific Instruments*, 84, 035006
- Kirz, J. & Jacobsen, C. (2009). *Journal of Physics: Conference series*, 186
- Kaulich, B., Thibault, P., Gianoncelli, A. and Kiskinova, M. (2011). *Journal of Physics: Condensed Matter*, 23, 083002
- Yan H., Conley R., Bouet N., and Chu Y. S. (2014). *J. Phys. D: Applied Physics*, in print.

Holt, M., Harder, R., Winarski, R. & Rose, V. (2013). *Annual Review of Materials Research*, 43, 183-211

Wu H. R. *et al.*, (2012). *J. Phys. D.*, 45 242001

Holler, M., Raabe, J., Diaz, A., Guizar-Sicairos, M., Quitmann, C., Menzel, A. and Bunk, O. (2012). *Review of Scientific Instruments*, 83, 073703

Holler, M., J., Diaz, A., Guizar-Sicairos, M., Karvine, P., Farm, E., Harkonen, E., Ritala, M., Menzel, A. Raabe, J. and Bunk, O. (2014). *Scientific Reports*, 4, 3857

Marsh, E. R. (2009). *Precision Spindle Metrology*, U. S. A.: DEStech Publications, Incorporated

Noire, P., Jonquieres, N., Schlutig, S., Leterme, D. and Roux., T. (2010). *Diamond Light Source Proceedings*, 1, 1-4

Macro, S. (2010). *Evaluation of the Sphere of Confusion of Goniometer Devices and Analysis of the SmarAct Linear Positioners*, internship report

Kim, J., Lauer, K., Yan, H., Chu, Y. S. and Nazaretski, E. (2013). *Review of Scientific Instruments*, 84, 035006

Figure 1 (a) Schematic view of the instrument. Major components are enumerated: (1) capacitive sensors; (2) high precision diamond turned aluminum reference cylinder; (3) X-Y piezo positioner; (4) rotary stage under testing. (b) Arrangement of five Lion Precision CPL 190 capacitive displacement sensors (S1-S5) used to map rotational errors. (c), (d) surface roughness and line profile of the aluminum reference cylinder acquired with Zygo white beam interferometer.

Figure 2 (a1-a2) Radial runout error of Stage 1 and Stage 2 (b1-b2) Wobble of Stage 1 and Stage 2 (c1-c2) Axial runout error of Stage 1 and Stage 2