REVISION HISTORY					
REV	DESCRIPTION	DATE	DESIGNER	CHECKED BY	ENGINEER
A	RELEASE PER ECO-004624	8/2017	M. JOHANSON	C. STELMACH	L. LIENHARD



NOTES:
1. REFERENCE DRAWING SR-FE-3PW-4001.
2. UNLESS OTHERWISE SPECIFIED THE 7-BM SOURCE POINT IS THE ORIGIN FOR ALL ORDINATE DIMENSIONS.
3. UNLESS OTHERWISE SPECIFIED, ORDINATE DIMENSIONS ARE TO THE U/S FACE FOR Pb SHIELDS AND Cu MASKS, TO THE GEOMETRIC CENTER FOR NON-OPTICAL COMPONENTS (E.G. Be WINDOW), AND TO THE OPTICAL CENTER FOR OPTICAL COMPONENTS.
4. SBT ENCLOSURE 1 MIN. LEAD THICKNESS: 5mm ALL AROUND.
5. SBT SHIELDED PIPE MIN. LEAD THICKNESS: 5mm WRAP ALL AROUND.
6. FOR COMPONENTS REFERENCED BY THIS NOTE SYNCHROTRON BEAM IMPINGES ON MASK AT NORMAL INCIDENCE OUTSIDE OF TAPERED APERTURE, PER HEAT LOAD CALCULATION "PD-QAS-MSK-CALC-1101" THIS IS ACCEPTABLE.

SPECIFICATIONS		REFERENCE
Project	BDN PROJECT	LT-EDP-015-042
Source	Three Pole Wiggler (3PW)	LT-C-SPC-SR-ID-BL-3PW-001 VERSION 1
Beam Line Name	QUICK X-RAY ABSORPTION SPECTROSCOPY BEAMLINE	LT-EDP-015-042
Abbreviation	QAS	LT-EDP-015-042
Period of insertion device	N/A	N/A
Cell	7-BM	LT-EDP-015-042
Length of Source	250 mm	LT-C-ASD-RSI-002
Location of Source (from Start of Bending Magnet)	366 mm	LT-C-ASD-RSI-002
Ray Tracing Procedure		PS-C-XFD-PRC-008 VERSION 2
Front End Ray Tracing Drawing		SR-FE-3PW-4001 REV B

ABBREVIATIONS:
BRS - BREMSSTRAHLUNG STOP
BS - BEAM STOP
CO - BREMSSTRAHLUNG COLLIMATOR
D/S - DOWNSTREAM
MONO - MONOCHROMATER
MSK - FIXED MASK
PBS - PINK BEAM STOP (FIXED MASK)
PCM - PHOTON COLLIMATING MIRROR
PFM - PHOTON FOCUSING MIRROR
PSH - PHOTON SHUTTER
RCO - RATCHET WALL COLLIMATOR
SBT - SHIELDED BEAM TRANSPORT
SCHM - SAMPLE CHAMBER
SLT - SLIT
U/S - UPSTREAM
WIN - WINDOW

QAS (7-BM) BEAMLINE LAYOUT
SCALE: 1:40

UNLESS OTHERWISE SPECIFIED
ALL DIMENSIONS ARE IN MILLIMETERS
DIMENSIONS IN BRACKETS [xx.xx] (WHERE PRESENT) ARE INCHES AND ARE FOR REFERENCE ONLY
INTERPRET DRAWING AS PER ANSI Y14.5-1994 OR Y32.3-1975

DIMENSIONAL TOLERANCES: X ± 1.500, .X ± 0.750, .XX ± 0.400, .XXX ± 0.120
ANGULAR TOLERANCE: ± .5°
FINISH: 3.2
BREAK EDGES & SHARP CORNERS 0.120 MIN. TO 0.750 MAX.

THIRD ANGLE PROJECTION
SCALE: SEE VIEW

BROOKHAVEN NATIONAL LABORATORY
BROOKHAVEN NATIONAL LABORATORY
BROOKHAVEN SCIENCE ASSOCIATES
UPTON, NEW YORK 11973
Exploring Life's Mysteries, Protecting its Future

NSLS II NATIONAL SYNCHROTRON LIGHT SOURCE II

DRAWN BY	M. JOHANSON	8/2017
CHECKED BY	C. STELMACH	8/2017
VACUUM APPROVAL	-	-
ENGINEER APPROVAL	L. LIENHARD	8/2017
SUPERVISOR APPROVAL	S. EHRLICH	8/2017
ES&H APPROVAL	L. STIEGLER	8/2017
QA APPROVAL	J. ZIPPER	8/2017

PHOTON DIVISION
QAS (7-BM)
BEAMLINE RAY TRACE

NEXT ASSY:	N/A	
PROJECT:	BDN	
WBS#	6.08.03	
ESH&Q RISK LEVEL	A1	
SHEET SIZE	E	
DRAWING/PART NUMBER	PD-QAS-RAYT-0001	
REVISION	A	
SHEET	1 OF 8	

C:\Mech-Eng\LTP\PD\QAS\RAYT\PD-QAS-RAYT-0001.dwg